



CANNON® II PLUS



NEXTSTEP® RETROGRADE



NEXTSTEP® ANTEGRADE



ARROW EDGE®

## ARROW® CHRONIC CATHETER

### Site-care guidelines

#### CATHETER CARE AND MAINTENANCE

Adhere to the following guidelines when caring for your Arrow Chronic Catheter:

- Alcohol, alcohol-based solutions (e.g. Hibiclens®, Chloraprep®), iodine-based solutions (e.g. Povidone-Iodine), PEG-based ointments (e.g. Bactroban®), hydrogen peroxide or ExSept Plus® are acceptable for use with these catheters.
- Ensure that solution is completely dry before applying an occlusive dressing.
- Avoid excessive or prolonged use of alcohol-based solutions and ointments to clean catheter or for site care.
- Do not use acetone with these catheters as catheter degradation may occur.
- Clean skin around catheter using accepted skin antiseptics. Cover exit site with sterile occlusive dressing for the entire duration of implantation. If swelling at the catheter site is observed, discontinue use and replace catheter.

#### HEPARINIZATION

To maintain patency of catheter between treatments, a heparin-lock must be created in each lumen of the catheter. Concentration of heparin used should be determined by hospital protocol. If catheter is not going to be used, it is recommended that catheter be flushed and re-packed with heparin every 48–72 hours.

#### MAGNETIC RESONANCE ENVIRONMENT

The NextStep Retrograde and Cannon II Plus are MR Conditional. Consult the Instructions For Use for further details.

TELEFLEX MEDICAL Research Triangle Park, NC 27709  
 Phone: 866.246.6990 Fax: 800.399.1028 Intl: 919.433.8088  
 TELEFLEX.COM

#### COMPATIBILITY OF SITE CARE ANTISEPTICS<sup>1</sup>

ANTISEPTIC AGENTS	CANNON II PLUS <sup>+</sup>	ARROW EDGE	NEXTSTEP ANTEGRADE	NEXTSTEP RETROGRADE
ExSept Plus	Yes	Yes	Yes	Yes
Povidone-Iodine <sup>2,3</sup>	Yes	Yes	Yes	Yes
Betadine	Yes	Yes	Yes	Yes
Tincture of Iodine	Yes	Yes	Yes	Yes
Alcohol-based solutions <sup>2,3</sup>	Yes	Yes	Yes	Yes
Chloraprep <sup>2,3</sup>	Yes	Yes	Yes	Yes
Isopropyl alcohol (IPA) <sup>2,3</sup>	Yes	Yes	Yes	Yes
Hibiclens	Yes	Yes	Yes	Yes
Hydrogen Peroxide	Yes	Yes	Yes	Yes
Bactroban <sup>2,3</sup>	Yes	Yes	Yes	Yes
Povidone-Iodine ointment <sup>2,3</sup>	Yes	Yes	Yes	Yes
PEG ointments	Yes	Yes	Yes	Yes
Acetone	No	No	No	No

<sup>1</sup> *In-vitro* test data available upon request.

<sup>2</sup> Centers for Disease Control and Prevention "Guidelines for the Prevention of Intravascular Catheter-Related Infections," August 9, 2002, Volume 51.

<sup>3</sup> National Kidney Foundation. NKF-K/DOQI Clinical Practice Guidelines for Vascular Access: update 2000. *AmJ Kidney Dis* 2001; 37 (sup): S137-S81.

<sup>+</sup> US Patent No. 6,638,242

Caution: U.S. Federal Law limits this device to sale by or on order of a physician.

Bactroban is a registered trademark of GlaxoSmithKline. Chloraprep is a registered trademark of CareFusion Corporation or one of its affiliates. ExSept Plus is a registered trademark of Alcavis HDC, LLC or one of its affiliates. Hibiclens is registered globally to one or more of the Mölnlycke Health Care Group of Companies. Arrow, Cannon, NextStep, Edge and Teleflex are trademarks or registered trademarks of Teleflex Incorporated or its affiliates. ©2012 Teleflex Incorporated. All rights reserved. 2012-0720

FOR MORE INFORMATION ABOUT THIS PRODUCT, VISIT [ARROWINTL.COM](http://ARROWINTL.COM)

TO PLACE AN ORDER CONTACT CUSTOMER SERVICE AT 1.866.246.6990